


LSR II Configuration



Note: There may be other compatible fluorophores/dyes that are not listed here. Always check fluorochrome combinations prior to running an experiment!!!

*Denotes filter change

¹7AAD and PI may be problematic depending on your fluorochrome panel.

²PerCP-Cy5.5 and PE-Cy5.5 cannot be combined.